



**Kenya**  
Red Cross



**CLARIFICATIONS ON THE TENDER FOR THE RE-TENDER OF THE SUPPLY AND DELIVERY OF STERILE NEEDLE AND SYRINGE (NSP KIT 1)**

**Q.** Syringe specific size. (NSP kit 1) specs information provided is insufficient for determining size.

The manufacturer has variance on available range size of syringes

**A.** The required needle gauge for the kit is 25Gx1/2

**Q.** The Insulin syringe specified is 3x1ml 25G, however the categories given by the manufacturer come with standard needle size:

29Gx1/2", or

30Gx5/16" and

31Gx15/64"

Please reaffirm on the given category

**A.** The Requirement is 25Gx1/2needle and 1 ml insulin syringe to which the gauge was informed by the users and is contained in the tender documents.

**Q.** 3 cotton filters (0.05microns). please clarify also on the size.

**A.** For the size of the cotton filters is 0.05 microns, kindly visit KRCS offices at Nairobi to view on the sample.

**Q.** We kindly request for a week extension of submission deadline date from 22nd to 29th June for **TENDER FOR THE SUPPLY AND DELIVERY OF STERILE NEEDLES AND SYRINGES KIT 1 (NSP KIT 1)**. This is after viewing the required samples at your stores following the pre bid meeting. We have been forced to change some items and need another shipment from our manufacturer.

**A.** This tender was already extended. This being a re-tender is already an additional time for applicants and with the timelines on the current stock, we are unable to extend it further.

Procurement Manager  
Kenya Red Cross Society

**ALL CORRESPONDENCE TO BE ADDRESSED TO THE SECRETARY GENERAL**

Kenya Red Cross Society, South "C" (Bellevue), Red Cross Road, Off Popo Road | P.O. Box 40712, 00100-GPO, Nairobi, Kenya.  
Tel: (+254-20) 6003593/6002465/3950000 | Fax: (+254-20) 3950444 | Mobile: (+254) 703 037000/722 206958/733 333040 email:  
info@redcross.or.ke | Website: www.redcross.or.ke

**Patron:** H.E. Hon. Uhuru Kenyatta CGH, President of the Republic of Kenya **Governor:** Mr. Francis Manthi Masika